



# Certificate of Instrument Performance

Company Name: CHELMSFORD WATER DISTRICT

Account Number: 133177

Contract Number: HACH1229836- C-Crooked Springs WTP

Certification Number: WO-01166175

Part Number: DR5000, oo nn rr DR5000 SPECTRO UV/VIS 115/230V	
Serial Number: 1190906	
Asset Tag : DR5000 SPEC	
<b>RECEIVED CONDITION:</b> <i>(One must be Checked)</i>	<input type="checkbox"/> Pre-Servicing Check Out Tests, <b>NOT</b> performed
	<input checked="" type="checkbox"/> Within Tolerance
	<input type="checkbox"/> Within Tolerance but Limited (*see servicing notes)
	<input type="checkbox"/> Out of Tolerance (*see servicing notes)
<b>RETURNED CONDITION:</b> <i>(One must be Checked)</i>	<input type="checkbox"/> Within Tolerance
	<input checked="" type="checkbox"/> Within Tolerance but Limited (*see servicing notes)
	<input type="checkbox"/> Out of Tolerance (*see servicing notes)
	Chemical Standards Used, (ID#):
Test Equipment Used, (ID#): Temp humidity pen i.d. 1099871 exp. 21Jun22; LZV 537 test filter set # 4924 exp. 31May23	
Environmental Conditions	
Temperature: 22 °C	Humidity: 69 %

*Hach Company does hereby certify that the above listed equipment meets or exceeds all Manufacturer's Service Specifications (unless limited conditions apply). Test equipment and chemicals used for performance verification are calibrated using standards traceable to the National Institute of Standards and Technology (NIST). Where such standards do not exist, the basis for calibration is documented. The proper operation of the above instrument was established at the time of certificate issuance. To insure continued performance, user must adhere to all requirements listed in the instrument manual.*

Certified by: Joseph Hodge

Signature:

Certification Date: 10/14/2021

Title: Authorized Service Representative